

Contact

JohnsonDiversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 12. 08. 2011

Product Name	CLEARKLENS DE VH29
Product Code	6069643
Batch Number	60 30 369
Production Date	07/2011
Expiry Date	07/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880 - 0.888	~ 0.885

"SIGNATURE"



Name A. GUMINSKA

Function Name PHARMA ROOM TEAM LEADER

Please note Johnsondiversey did change the company name to Diversey and is in the phase of implementation

FLEDD1

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Isotron Limited

Roydsdale Way · Euoway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

LA1 4XS
 P017305
 Cust. Ref: P017305
 Date Rec'd 4/08/11
 Date: 10/08/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLEDD10121 B211080224	JD DE 900ML BOTTLES 25-45kGy	699	172 X B/No.W030369 CUST-B/No.2723 527 X B/No.W030548 CUST-B/No.2863 ACTUAL DOSE....36.9 - 40.2....kGy
B2FLEDD10999 B211080224	SAMPLES	14	9 X B/No.W030369 5 X B/No.W030548 CUSTOMER BATCH/CODE No.s AS ORDER No.P017305 DATED 3/8/11
Despatch Text:			
14PLTS .			
Total		713	Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature: *J. Coffey*
 For and on behalf of ISOTRON LIMITED



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 15 Aug 11
Date Tested: 18 Aug 11
Date Completed: 18 Aug 11
Purchase Order: -

Test Article: ClearKlens DE Block 1
Sample Code: WO30369
Batch Ref: 2723
Qty Received: 1 x 900mL Bottle
Laboratory Reference: 0002181/1
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	99 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	MM110/05	FM01

External Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 19 Aug 11 15:35:52

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0002181
Lancaster

Certificate of Analysis - OSMM

Print Number

P0004736

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited



CERTIFICATE OF ANALYSIS

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 15 Aug 11
Date Tested: 19 Aug 11
Date Completed: 02 Sep 11
Purchase Order: -

Test Article: ClearKlens DE Block 1
Sample Code: WO30369
Batch Ref: 2723
Qty Received: 20 x 900mL + 3 Spare Bottles
Laboratory Reference: 0002180/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Product Standard Data Sheet	MM107/03	FM03
Volume Tested	MM107/05	20 x 50 mL

External Comments

Test carried out according to current Pharmacopocias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 08 Sep 11 11:00:27

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0002180
Lancaster

Certificate of Analysis - OSMM

Print Number

P0004937

This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited