

**Contact**

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: 29/11/11

<b>Product Name</b>	CLEARKLENS TEGO 2000SC VH25
<b>Product Code</b>	7513387
<b>Batch Number</b>	WJ 30230
<b>Production Date</b>	11/2011
<b>Expiry Date</b>	11/2013

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Slightly Yellow Liquid	Clear Slightly Yellow Liquid
pH 100% (20°C)	JMD001	7.5 - 8.5	7.8
Specific Gravity (20°C)	JMD004	0.993 - 1.003	0.997

"SIGNATURE" *C. Pliscoe*

Name *Caroline Pliscoe*

Function Name *Plantroom Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit 8  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**

Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P017791  
 Date Rec'd 17/11/11  
 Date: 22/11/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B211110684	DIVERSEY TEGO 2000SC 50ml 25-42kGy	105	B/No.M030830 - CUSTOMER B/No.379148 ACTUAL DOSE....32.3 - 33.6....kGy
B2FLE0010999 B211110684	SAMPLES	2	CUSTOMER BATCH/CODE No.s AS ORDER No.P017791 DATED 16/11/11
Despatch Text: 2PLTS			
Total		107	Last Page 1 of 1

All in accordance with current  
 Technical Agreement

This is to certify that the above items have been irradiated as  
 specified above

Authorising Signature: *J. Goffey*

For and on behalf of ISOTRON LIMITED



**CERTIFICATE OF ANALYSIS**

Mr J Jackson  
Flexible Medical Packaging Limited  
Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

**Date Received:** 01 Dec 11  
**Date Tested:** 07 Dec 11  
**Date Completed:** 29 Dec 11  
**Purchase Order:** 17858

**Test Article:** Clarklens Tego 2000 SC VII25  
**Sample Code:** WO30830  
**Batch Ref:** 379148  
**Qty Received:** 20 x 50mL  
**Laboratory Reference:** 0002885/1  
**Test Required:** Sterility Test by Membrane Filtration (Steritest) Method I with Subculture: PH1 Eur and USP

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

Approval is provided by Electronic Signature. Their name is shown below.

Date: 05 Jan 12 10:07:46

Mrs S.L. Wood, BSc (Hons), CBiol, MSB  
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0002885  
Lancaster

Certificate of Analysis - OSMM

Print Number

P0001089

*This report shall not be reproduced except in full, without the written approval of Wickham Laboratories Limited*