

FLE001

Page 1

Flexible Medical Packaging Ltd  
 Unit B  
 White Cross Industrial Estate  
 Hightown  
 Lancaster



**Isotron Limited**  
 Roydsdale Way · Euroway Trading Estate  
 Bradford · West Yorkshire  
 BD4 6SE · United Kingdom  
 Telephone: +44 (0)1274 686011  
 Facsimile: +44 (0)1274 686061  
 Email: assistance@isotron.com  
 Website: www.isotron.com

LA1 4XS

Cust. Ref: P017686  
 Date Rec'd 27/10/11  
 Date: 4/11/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B211110013	JD DE 900ML BOTTLES 25-45kGy	355	B/No.W031000 - CUSTOMER B/No.3020  ACTUAL DOSE....37.2 - 38.4....kGy
B2FLE0010999 ...11110013	SAMPLES	1	CUSTOMER BATCH/CODE No.s AS ORDER No.P017686 DATED 26/10/11
Despatch Text:  7PLTS			
Total		356	Last Page 1 of 1

All in accordance with current  
 Technical Agreement

This is to certify that the above items have been irradiated as  
 specified above

Authorising Signature: *T. Coffey*

For and on behalf of ISOTRON LIMITED



**Contact**

JohnsonDiversey Europe BV  
Maarssebroeksedijk 2  
3542 DN Utrecht  
The Netherlands  
Tel. +31 (0)30 247 6427

**CERTIFICATE OF ANALYSIS**

Date: *5/11/11*

<b>Product Name</b>	CLEARKLENS DE STANDARD VH29
<b>Product Code</b>	7515499
<b>Batch Number</b>	<i>no 3/000</i>
<b>Production Date</b>	<i>10/2011</i>
<b>Expiry Date</b>	<i>10/2013</i>

This is to certify that the above batch of product has been tested internally and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits	Results
Appearance	Visual	Clear Colourless Liquid	<i>clear colourless liquid</i>
Specific Gravity (20°C)	JDM004	0.880 - 0.888	<i>0.885</i>

"SIGNATURE" *C. Ascoe*

Name *Caroline Ascoe*  
Function Name *Pharmacy Supervisor*

Please note JohnsonDiversey did change the company name to Diversey and is in the phase of implementation