



Diversey Europe BV
 Maarssebroeksedijk 2
 3542 DN Utrecht
 The Netherlands
 Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 20/1/12

Product Name	CLEARKLENS DE STANDARD VH29
Product Code	7515499
Batch Number	W031293
Production Date	01/2012
Expiration Date	01/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880	0.888	0.885

On behalf of Diversey site Quality Manager	Name :	<i>C. Pistor</i>
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7515499	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Synergy Health plc
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@synergyhealthplc.com
 Website: www.synergyhealthplc.com

LA1 4XS

Cust. Ref: P018001
 Date Rec'd 12/01/12
 Date: 16/01/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B212010264	JD DE 900ML BOTTLES 25-45kGy	712	96 X B/No.W031326 CUSTOMER B/No.3184 616X B/No.W031293 CUSTOMER B/No.3184 ROUTINE DOSE 35.8, 36.1, 38.3, 38.9 kGy
B2FLE0010999 B212010264	SAMPLES	2	CUSTOMER BATCH/CODE No.s AS ORDER No.P018001 DATED 11/1/12
Despatch Text: 14PLTS			
Total		714	Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature: *J. Gifford*

For and on behalf of SYNERGY HEALTH PLC