



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 26/3/12

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	7513387
Batch Number	W031920
Production Date	03/2012
Expiration Date	03/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow	clear slightly yellow liquid
pH 100% (20°C)	JDM001	7.5	8.5	7.6
Specific Gravity (20°C)	JDM004	0.993	1.003	0.997

On behalf of Diversey site Quality Manager	Name :	<i>C. P. Scaer</i>
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513387	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

CERTIFICATE OF IRRADIATION

Number: **B**

0524501 00

Page 1

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@synergyhealthplc.com
 Website: www.synergyhealthplc.com

LA1 4XS
 P018310
 Cust. Ref: 15/03/12
 Date Rec'd 19/03/12
 Date:

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010094 B212030514	DIVERSEY TEGO 2000SC 50ml 25-42kGy	303	236 X B/No.W031488 67 X B/No.W031920 ROUTINE DOSE 34.3, 34.5, 35.5 kGy
B2FLE0010999 B212030514	SAMPLES	4	1 X B/No.W031488 3 X B/No.W031920 CUSTOMER BATCH/CODE No.s AS ORDER No.P018310 DATED 14/3/12
Despatch Text: 7PLTS		Total	307
			Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC



Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 21 Mar 2012
Date Tested: 27 Mar 2012
Date Test Completed: 18 Apr 2012
Purchase Order: PO18328

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0004020/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I with Subculture: PH Eur and USP
Date Received: 21/03/2012
Test Article: Clearklens Tego 2000 SC VH25
Sample Code:
Batch Ref: WO31920
Qty Received: 20 x 50mL

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Growth of Sub-culture in Tryptone Soya Broth at 30-35°C after 4 days	MM107/07	No growth in two broths
Growth of Sub-culture in Fluid Thioglycollate at 20-25°C after 4 days	MM107/08	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM08

Comments

As Steritest Sterility Test

Approval is provided by Electronic Signature. Their name is shown below.

Date: 23 Apr 2012 09:16:47

Mrs P. Pham-Lengoc, BSc, MSB
Laboratory Manager - Pharmaceutical Microbiology