



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands
Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 4/5/12

Product Name	CLEARKLENS DE VH29
Product Code	7515783
Batch Number	W231943
Production Date	04/2012
Expiration Date	04/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear colourless liquid
Specific Gravity (20°C)	JDM004	0.880	0.888	0.881

On behalf of Diversey site Quality Manager	Name :	<i>C. P. van der Vliet</i>
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7515783	Version : 03	Date of issuing : November 28 th 2011
-------------	--------------	--

FLE001

Page 1

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Synergy Health plc
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@synergyhealthplc.com
 Website: www.synergyhealthplc.com

LA1 4XS
 P018486
 Cust. Ref: 26/04/12
 Date Rec'd: 30/04/12
 Date:

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010121 B212040847	JD DE 900ML BOTTLES 25-45kGy	172	B/No.W031943 - PRODUCT REF.7515783 ROUTINE DOSE 37.9, 36.8, 39.2 kGy
B2FLE0010999 B212040847	SAMPLES	5	 CUSTOMER BATCH/CODE No.s AS ORDER No.P018486 DATED 25/4/12
	Despatch Text: 4PLTS		
	Total	177	Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature: *J. (off)*

For and on behalf of SYNERGY HEALTH PLC

Mr J Jackson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 08 May 2012
Date Tested: 10 May 2012
Date Test Completed: 10 May 2012
Purchase Order: 18542

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0004554/2
Test Required: Bacterial Endotoxin Testing by Kinetic QCL: Ph Eur and USP
Date Received: 08/05/2012
Test Article: Clearklens DE Block 2
Sample Code: 3442
Batch Ref: WO 31943
Qty Received: 1 x 900mL

Test	Method Item	Result
Bacterial Endotoxin by Kinetic QCL Test	MM110/00	<0.25 EU/mL
Spiked Recovery	MM110/01	95 %
Dilution Tested	MM110/04	1/50
Endotoxin Spike Level	MM110/02	0.5 EU/mL
Product Standard Data Sheet	FG047/psd	FM01

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.



Date: 14 May 2012 17:19:37

Mrs S.L. Wood, BSc (Hons), CBiol, MSB
Senior Laboratory Manager - Pharmaceutical Microbiology

Consignment: 0004554

Certificate of Analysis - OSMM

Print Number: P0002989

Page 1 of 1



Mr J Jackson
 Flexible Medical Packaging Limited
 Unit 8
 Hightown
 White Cross Industrial Estate
 Lancaster
 LA1 4XS

Date Received: 08 May 2012
Date Tested: 15 May 2012
Date Test Completed: 29 May 2012
Purchase Order: 18542

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0005048/1
Test Required: Sterility Test by Membrane Filtration (Steritest) Method I: Ph Eur and USP
Date Received: 08/05/2012
Test Article: ClearKlens DE Block 2
Sample Code: 3442
Batch Ref: WO 31943
Qty Received: 20 x 900mL + Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in one broth
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in one broth
Volume Tested	MM107/05	20 x 50 mL
Product Standard Data Sheet	FG047/psd	FM03

Comments

Test carried out according to current Pharmacopoeias.

Approval is provided by Electronic Signature. Their name is shown below.

Date: 06 Jun 2012 14:20:02

Mrs P. Pham-Lengoc, BSc, MSB
 Laboratory Manager - Pharmaceutical Microbiology