



Diversey Europe BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427

CERTIFICATE OF ANALYSIS

Date: 27/6/12

Product Name	CLEARLEN S TEGO 2001SC VH24
Product Code	7517177
Batch Number	W032435
Production Date	06/2012
Expiration Date	06/2014

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow	clear slightly yellow liquid
pH 100% (20°C)	JDM001	8.5	9.5	8.7
Specific Gravity (20°C)	JDM004	1.040	1.080	1.053

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7517177	Version : 02	Date of issuing : November 28 th 2011
-------------	--------------	--

FLE001

Flexible Medical Packaging Ltd
 Unit 8
 White Cross Industrial Estate
 Hightown
 Lancaster



Synergy Health plc

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@synergyhealthplc.com
 Website: www.synergyhealthplc.com

LA1 4XS
 P018725
 Cust. Ref: 21/06/12
 Date Rec'd
 Date: 25/06/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2FLE0010091 B212060706	DIVERSEY TEG02001 10x5x100 25-42kGy	10	B/No.W032435 - CUSTOMER B/N.395932 ROUTINE DOSE 32.3 kGy
B2FLE0010999 B212060706	SAMPLES	2	CUSTOMER BATCH/CODE No.s AS ORDER No.P018725 20/6/12
Despatch Text: 1PLT			
Total		12	Last Page 1 of 1

All in accordance with current
 Technical Agreement

This is to certify that the above items have been irradiated as
 specified above

Authorising Signature:

For and on behalf of SYNERGY HEALTH PLC