



Diversey Europe Operations BV

Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427


CERTIFICATE OF ANALYSIS

Date: 15/2/16

Product Name	CLEARLEN S TEGO 2000SC VH25
Product Code	100868202
Batch Number	FMP16027 39704
Production Date	27/01/2016
Expiration Date	EXP 27/01/2019

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear slightly yellow liquid
pH 100% (20°C)	JDM001	7.5	8.5	7.8
Specific Gravity (20°C)	JDM004	0.993	1.003	0.993

On behalf of Diversey site Quality Manager	Name :	
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA 7513387	Version : 04	Date of issuing : April 27 th 2012
-------------	--------------	---



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 07-Feb-2016

UK32S11558253-1-1

This is to certify that Bradford Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Flexible Medical Packaging Ltd
 Unit 8, White Cross Ind Estate
 Hightown
 Lancaster
 Lancashire LA1 4XS
 UNITED KINGDOM

All in accordance with current Technical Agreement

Order Information

Account Number:	100432
Synergy Health Sales Part Reference:	1002722
Customer Reference Number:	P025924
Product Description:	DIVERSEY TEGO 2000SC 50ml 25-42kGy
Validation Reference:	4.1970
Quantity Received:	133
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	42.0
Customer Unit Lot/Batch Number:	103 X B/N: FMP16027 39704 +1 SAMPLE, 28 X B/N: FMP16032 39802 +1 SAMPLE THU 04 FEB 2016 3PLTS LOT 1

Irradiation Data

Date and Time of Irradiation:	07-Feb-2016 08:26
Reference Dose Range kGy:	33.6 - 34.6
Calculated Minimum Dose kGy:	27.2
Calculated Maximum Dose kGy:	38.2

Irradiation Release Authorised By Synergy Health plc

Processing Site: Roysdale Way, Euroway Trading Estate, Bradford, BD4 6SE Phone No: +44 (0) 1274 686011

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered In England and Wales No: 01771333 VAT Number: GB 813038069



Wickham Laboratories

Contract Analytical Services

Wickham Laboratories Ltd
Hoeford Point, Barwell Lane, Gosport
Hampshire PO13 0AU England
Telephone: +44(0)1329 226600
Fax: +44(0)1329 226688

mail@wickhamlabs.co.uk
www.wickhamlabs.co.uk

Ms D Henderson
Flexible Medical Packaging Limited
Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

Date Received: 15 Feb 2016
Date Tested: 18 Feb 2016
Date Test Completed: 03 Mar 2016
Purchase Order: 25984

CERTIFICATE OF ANALYSIS

Laboratory Reference Number: 0027292/2
Test Required: Sterility by Membrane Filtration Steritest
Date Received: 15/02/2016
Test Article: ClearKlens TEGO 2000sc
Sample Code: FMP16027 39704
Batch Ref: 518692
Qty Received: 20 x 50mL Bottles + 5 Spare

Test	Method Item	Result
Sterility Test by Membrane Filtration (Steritest) Method	MM107/00	Pass
Growth in Tryptone Soya Broth at 20-25°C after 14 days	MM107/01	No growth in two broths
Growth in Fluid Thioglycollate Medium at 30-35°C after 14 days	MM107/02	No growth in two broths
Volume Tested	MM107/05	20 x 50 mL
Tested in Accordance with Ph Eur, USP & JP	MQ005/07	EP 8.0 2.6.1, USP 38 <71> & JP XVI 4.06
Product Standard Data Sheet	FG047/psd	FM08

Approval is provided by Electronic Signature. Their name and position is shown below.

Date: 04 Mar 2016 14:12:44

Mrs P. Pham-Lengoc
Laboratory Manager - Pharmaceutical Microbiology