



Diversey Europe Operations BV  
Maarssebroekse dijk 2  
3542 DN Utrecht  
The Netherlands

Tel. +31 (0)30 247 6427  
www.Diversey.com

## CERTIFICATE OF ANALYSIS

Date: 11.11.2019

Product Name	CLEARKLENS TEGO 2000SC VH25
Product Code	100961418
Batch Number	FMP19277, 48470
Production Date	04/10/2019
Expiration Date	EXP 04/10/2022

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	clear slightly yellow liquid
Specific Gravity (20°C)	DM004	0.993	1.003	1.002
pH (neat solution)	DM001	7.5	8.5	7.8

On behalf of Diversey site Quality Manager	Name :	Navabunda Pantyus
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24 <sup>th</sup> 2017
--------------	--------------	--

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-5756A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
TEGO 2000SC DV4724	FMP19277, 48470	98	Case
Validation Reference Number: 4724			

Processing Run Start Date: 22-Oct-19 12:27 PM

Processing Run End Date: 22-Oct-19 07:33 PM

Specified Dose Range (kGy):	25.0 - 45.0	Calculated Min Dose (kGy):	28.8
Reference Dose Range (kGy):	29.4- 40.5	Calculated Max Dose (kGy):	38.0

Gamma Process Run Approval authorized by STERIS

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK  
Limited, a STERIS Company  
Brunel Close  
Drayton Fields Industrial Estate  
Daventry  
Northants  
NN11 8RB  
Phone: + 44(0) 1327 706 111

# PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Tego 2000 SC

**Sample Description:** Sample Code: Fmp19277 48470

**Lucideon Sample Number:** (196202)-47210

**Lucideon Report Number:** (196202)-47210/MFEP

**Issue Number:** 1

**Date Logged:** 12-Nov-2019

**Order Number:** 34274

**Date Reported:** 03-Dec-2019

**Date(s) of Test(s):** 16-Nov-2019 to 30-Nov-2019

### Sterility Testing

Membrane Filtration EP

## Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass

Tests carried out in accordance with cGMP

**End of Test Report**

Natalie Boot 03-Dec-19

Mrs Natalie Boot

**Senior Business Support Assistant**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com