



Diversey Europe BV
Maarssebroekse dijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 12/11/2020

Product Name	CLEARKLENS CLEANSINALD SS VH9S
Product Code	7516431
Batch Number	FMP 20317 50343
Production Date	12/11/2020
Expiration Date	EXP 12/05/2022

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear	Slightly Yellow Liquid	Clear slightly yellow liquid
pH (neat solution)	DM001	9.0	11.4	10.8
Specific Gravity (20°C)	DM004	0.990	1.010	1.001

On behalf of Diversey site Quality Manager	Name :	A. Bony, Harabinda
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of issuing : November 24 th 2017
--------------	--------------	--

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-13872A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
C/SINALD SS 900ml Bottle DV4673	WO50342	13	Case
Validation Reference Number: 4673			

Processing Run Start Date: 03-Oct-20 02:40 AM

Processing Run End Date: 03-Oct-20 08:10 AM

Specified Dose Range (kGy):	25.0 - 40.0	Calculated Min Dose (kGy):	29.1
Reference Dose Range (kGy):	27.3 - 36.6	Calculated Max Dose (kGy):	35.9

PO Number: 36453

Gamma Process Run Approval authorized by STERIS

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location
Synergy Health Sterilisation UK
Limited, a STERIS Company
Brunel Close
Drayton Fields Industrial Estate
Daventry
Northants
NN11 8RB
Phone: + 44(0) 1327 706 111

Items Irradiated under WO50342 will be used in finished batch Fmp20317 50343.

Olga Kirchner

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Cleansinald SS

Sample Description: Sample Code: Fmp20317 50343

Lucideon Sample Number: UK204707-33229

Lucideon Report Number: UK204707-33229/MFEP **Issue Number:** 1

Date Logged: 18-Nov-2020 **Order Number:** PO 36848

Date Reported: 11-Dec-2020 **Date(s) of Test(s):** 27-Nov-2020 to 11-Dec-2020

Sterility Testing

Membrane Filtration EP


Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass

Tests carried out in accordance with cGMP

End of Test Report


11 DEC 20

Mrs Andrea Saunders

Business Support Assistant

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com