

Diversey Europe Operations EV Masrasenbrooksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

## **CERTIFICATE OF ANALYSIS**

Date:

06 January 2021

Product Name	CLEARKLENS IPA VH1			
Product Code	101100418			
Batch Number	FMP20350	50728		
Production Date	15/12/2020			
Expiration Date	EXP 15/12/2022			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid	
Specific Gravity (20°C)	DM004	0.872	0.883	0.879	

On behalf of Diversey site Quality Manger	Name:	Urszula Haraburda Angelika Partynska
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



## **Certificate of Processing**

# STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

12/29/2020 10:07:23 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2798769
P.O.# PO37061 Sales Order # 2661867

**25 - 45 kGy** Gamma Irradiation Irradiation Date/Time: 12/27/2020 18:29:24 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	708 BX	101100418	Diversey Clearklens IPA/DE 900	FMP20350	WO50728/DEV002080
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
102.000	12 BX	101100418_SAMPLE	Diversey Clearklens IPA/DE 900	FMP20350	WO50728/DEV002080
	Dose Map	302_0030	Ref Pos: 10E1, AF Min = 0.71		
	720 BX	Total			

#### **Quality Test Summary**

		, · · · · · · · · · · · · · · · · ·			Signed By	_	
Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.7 KGY	Pass	IWEATHERAL IAN WEATHERALL	12/28/2020 17:07:18 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	40.0 KGY	Pass	IWEATHERAL IAN WEATHERALL	12/28/2020 17:07:39 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ASHLEY HOBSON Date: 12/29/2020 10:04:11 GMT

Reason: Work Order Completions

## PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA1 4XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20350 50728

Lucideon Sample Number: UK2166-611

Lucideon Report Number: UK2166-611/MFEP

Issue Number:

Date Logged:

08-Jan-2021

Order Number:

PO 37156

Date Reported:

Date(s) of Test(s):

22-Jan-2021 to 05-Feb-2021

08-Feb-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report** 



Mrs Natalie Boot

**Business Support Administrator** 

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.

## PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp20350 50728

Lucideon Sample Number: UK2166-612

Lucideon Report Number: UK2166-612/ETEP

Issue Number:

Date Logged:

08-Jan-2021

Order Number:

PO 37156

Date Reported:

Date(s) of Test(s):

11-Feb-2021

21-Jan-2021 to 21-Jan-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2EU/ml)

Tests carried out in accordance with cGMP

**End of Test Report** 

0+11-Reb-21

Mrs Natalie Boot

**Business Support Administrator**