

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 17/0/2021

Product Name	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26
Product Code	7514852
Batch Number	FMP21048; 51124
Production Date	17/02/2021
Expiration Date	EXP 17/02/2023

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Lin Lower	nits Upper	Results
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
pH (neat solution)	DM001	9.0	12.5	11.5
Specific Gravity (20°C)	DM004	1.004	1.020	1,005
Potential Chlorine Dioxide (ppm)		100	180	141.75

On behalf of Diversey site	Name:	Vavalende Partyusto
Quality Manager	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version: 02	Date of Issuing: November 24th 2017
--------------	-------------	-------------------------------------



Diversey Europe BV Maarseenbroeksedlijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 17/02/2021

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26
Product Code	7514852
Batch Number	FMP21048; 51124
Production Date	17/02/2021
Expiration Date	EXP 17/02/2023

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Y	nits · Upper	Results
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	1.5	2.5	2.0
Specific Gravity (20°C)	DM004	1.010	1.030	1.016

On behalf of Diversey site Quality Manager	Name :	Howalinda Partyust
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of Issuing: November 24th 2017
--------------	--------------	-------------------------------------

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-15904A

Product CodeLot NumberQuantityUOMBI-SP B3 COMPONENTS DV4725WO511238Case

Validation Reference Number: 4725

Processing Run Start Date: 23-Dec-20 07:47 PM

Processing Run End Date: 24-Dec-20 02:00 AM

Specified Dose Range (kGy): 25.0 - 40.0 Calculated Min Dose (kGy): 26.8

Reference Dose Range (kGy): 30.2 - 39.2 Calculated Max Dose (kGy): 33.7

PO Number: 37043

Gamma Process Run Approval authorized by STERIS

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK Limited, a STERIS Company Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB

Phone: + 44(0) 1327 706 111

Items irradiated under WO51123 will be used in finished batch FMP21048 51124.

Olga Kirchner

Rel Date: 08/13/2018

Document ID: 68147

Last Revised in Rel 2.0.0.0

Page 1 of 1

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO: Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: Fmp21048 51124

Lucideon Sample Number: UK21859-5873

Lucideon Report Number: UK21859-5873/MFEP

Issue Number:

Date Logged:

26-Feb-2021

Order Number:

PO37550

Date Reported:

19-Mar-2021

Date(s) of Test(s):

04-Mar-2021 to 18-Mar-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

MB00+ 19-401-21

Mrs Natalie Boot

Business Support Administrator

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: Fmp21048 51124

Lucideon Sample Number: UK21859-5874

Lucideon Report Number: UK21859-5874/MFEP

Issue Number:

Date Logged:

26-Feb-2021

Order Number:

PO37550

04-Mar-2021 to 18-Mar-2021

Date Reported:

19-Mar-2021

Date(s) of Test(s):

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

xBOOK A-MOr-21

Mrs Natalie Boot

Business Support Administrator