

Diversey Europe Operations BV Maarseenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

CERTIFICATE OF ANALYSIS

Date:

23 June 2021

Product Name	CLEARKLENS IPA VH1				
Product Code	7513400				
Batch Number	FMP21140	51841			
Production Date	20/05/2021				
Expiration Date	EXP 20/05/2023				

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	its - Upper	Results	
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid	
Specific Gravity (20°C)	DM004	0.872	0.883	0.875	

On behalf of Diversey site	Name:	Justyna Staron	Angelika P <mark>artyns</mark> ka
Quality Manger	Position	Quality	Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 02 Date of issuing: November 24th



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102

06/14/2021 11:02:04 GMT

Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2895110 P.O.# PO38154 Sales Order # 2781297

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 06/11/2021 05:24:49 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	398 BX	7513400	Diversey Clearklens IPA 900 mL	FMP21137	WO51840/DEV001152
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP21137	WO51840/DEV001152
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
103.000	549 BX	7513400	Diversey Clearklens IPA 900 mL	FMP21140	WO51841/DEV001152
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
	950 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	- Date /Time
450.00	Minimum Dose	25.0 kGv	45.0 kGv	27.7 KGY	Pass	JYATES	06/11/2021 17:18:37 GMT
430.00	Willimum Dose	Reason Code Test		27.7 KG1	r a 5 5	JOHN YATES	00/11/2021 17.10.37 GWII
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.1 KGY	Pass	JYATES	06/11/2021 17:18:52 GMT
		Reason Code Test				JOHN YATES	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 06/14/2021 11:01:54 GMT

Reason: Work Order Completions



Certificate of Processing

STERIGENICS Seymour Link Road Woodthorpe Mastin Moor TEL FAX www.sterigenics.com

R55480102 06

06/21/2021 12:28:57 GMT Page 1 of 1

Customer Name: FLEXIBLE MED. GBP UK LANCASTER Processing Facility: Markham Vale Work Order # 2906966 P.O.# 2795675

25 - 45 kGy Gamma Irradiation Irradiation Date/Time: 06/19/2021 01:32:35 GMT

SO Line #	Qty UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	878 BX	7513400	Diversey Clearklens IPA 900 mL	FMP21140	WO51841/STR001152
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
102.000	3 BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP21140	WO51841/STR001152
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
103.000	69 BX	101101081	Diversey Clearklens DE 900ml	FMP21146	WO51895/STR002096
	Dose Map	302_0112	REF: 13.5A3 AF Min = 0.717		
	950 BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	User	Date /Time
450.00	Minimum Dose	25.0 kGy Reason Code Test	45.0 kGy	27.6 KGY	Pass	JBEARNE JARETH BEARNE	06/19/2021 16:34:06 GMT
450.00	Maximum Dose	25.0 kGy Reason Code Test	45.0 kGy	41.0 KGY	Pass	JBEARNE JARETH BEARNE	06/19/2021 16:34:26 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS Date: 06/21/2021 12:28:39 GMT

Reason: Work Order Completions

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp21140 51841

Lucideon Sample Number: UK212768-18916

Lucideon Report Number: UK212768-18916/ETEP

Issue Number:

Date Logged:

25-Jun-2021

Order Number:

400144

Date Reported:

Date(s) of Test(s):

06-Jul-2021

to 06-Jul-2021

08-Jul-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

NCBOOK 08-JUI-21

Mrs Natalie Boot

Business Support Administrator

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested.

No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation

PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Clear Klens IPA

Sample Description:

Sample Code: Fmp21140 51841

Lucideon Sample Number: UK212768-18915

Lucideon Report Number: UK212768-18915/MFEP

Issue Number:

Date Logged:

25-Jun-2021

Order Number:

400144

28-Jun-2021 to 12-Jul-2021

Date Reported:

14-Jul-2021

Date(s) of Test(s):

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report



Mrs Natalie Boot

Business Support Administrator