

## CERTIFICATE OF ANALYSIS

Date Issue: 20 July 2021


<b>Product Name</b>	DI ClearKlens IPA VH1 S 6x0.9L
<b>Product Code</b>	<input checked="" type="checkbox"/> 101105062/ <input type="checkbox"/> 101105063/ <input type="checkbox"/> 101105064
<b>Batch Number</b>	FMP21175 52072
<b>Production date</b>	24/06/2021
<b>Expiration Date</b>	EXP 24/06/2023

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Specification	Unit	Results
Appearance	Visual	Colourless Clear Liquid	N/A	Colourless Clear Liquid
Specific Gravity (20°C)	DM 004	0.872 - 0.883	g/cm <sup>3</sup>	0.875

Control of the Sterility of the product:

Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 - 45.0 (kGy)	Yes	52072
Sterility certificate	Sterility by Membrane Filtration EP 9.0 2.6.1, USP 42<71> & JP XVII 4.06	No Growth	Yes	UK213134-21259
2.6.14 Bacterial Endotoxins	Method C - Turbidimetric Kinetic Method (EP/USP Criteria)	<0.2 EU/mL	Yes	UK213134-21260

	Date of released On behalf of Diversey's Manufacturing location: 20/07/2021	by : <i>CBoer</i>
		Position: <i>QA ADMIN</i>

The analysis results above could change over time and in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the SDS.

\*\*\*End of certificate of analysis\*\*\*

R55480102

08/13/2021 10:41:09 GMT  
Page 1 of 1

<b>Customer Name:</b>	<b>FLEXIBLE MED. GBP UK LANCASTER</b>	<b>Processing Facility:</b>	<b>Markham Vale</b>	<b>Work Order #</b>	<b>2916004</b>
<b>P.O.#</b>	<b>400148</b>			<b>Sales Order #</b>	<b>2806512</b>
	<b>25 - 45 kGy</b>	<b>Gamma Irradiation</b>	<b>Irradiation Date/Time:</b>		<b>07/03/2021 08:53:08 GMT</b>

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
106.000	824	BX	101105062	Diversey Clearklens IPA 900ml	FMP21169	WO52071/STR002199
			302_0112	REF: 13.5A3 AF Min = 0.717		
108.000	3	BX	101105062_SAMPLE	Diversey Clearklens IPA 900ml	FMP21169	WO52071/STR002199
			302_0112	REF: 13.5A3 AF Min = 0.717		
109.000	123	BX	101105062	Diversey Clearklens IPA 900ml	FMP21175	WO52072/STR002199
			302_0112	REF: 13.5A3 AF Min = 0.717		
	<b>950</b>	<b>BX</b>	<b>Total</b>			

**Quality Test Summary**

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.0 KGY	Pass	DMARRIOTT	07/03/2021 16:56:03 GMT
		Reason Code Test				DAVID MARRIOTT	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.1 KGY	Pass	DMARRIOTT	07/03/2021 16:56:22 GMT
		Reason Code Test				DAVID MARRIOTT	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS  
Reason: Work Order Completions

Date: 08/13/2021 10:40:58 GMT

R55480102

07/12/2021 10:25:06 GMT  
Page 1 of 1

<b>Customer Name:</b>	<b>FLEXIBLE MED. GBP UK LANCASTER</b>	<b>Processing Facility:</b>	<b>Markham Vale</b>	<b>Work Order #</b>	<b>2922102</b>
<b>P.O.#</b>	<b>400206</b>			<b>Sales Order #</b>	<b>2813946</b>
	<b>25 - 45 kGy</b>	<b>Gamma Irradiation</b>	<b>Irradiation Date/Time:</b>		<b>07/09/2021 23:08:47 GMT</b>

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
105.000	950	BX	101105062	Diversey Clearklens IPA 900ml	FMP21175	WO52072/STR002199
			Dose Map	REF: 13.5A3 AF Min = 0.717		
	<b>950</b>	<b>BX</b>	<b>Total</b>			

**Quality Test Summary**

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By-----	
						User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.3 KGY	Pass	IWEATHERAL	07/10/2021 16:17:10 GMT
		Reason Code Test				IAN WEATHERALL	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.1 KGY	Pass	IWEATHERAL	07/10/2021 16:17:32 GMT
		Reason Code Test				IAN WEATHERALL	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS  
Reason: Work Order Completions

Date: 07/12/2021 10:24:53 GMT

R55480102

07/12/2021 10:38:50 GMT  
Page 1 of 1

<b>Customer Name:</b>	<b>FLEXIBLE MED. GBP UK LANCASTER</b>	<b>Processing Facility:</b>	<b>Markham Vale</b>	<b>Work Order #</b>	<b>2922104</b>
<b>P.O.#</b>	<b>400209</b>			<b>Sales Order #</b>	<b>2813948</b>
	<b>25 - 45 kGy</b>	<b>Gamma Irradiation</b>	<b>Irradiation Date/Time:</b>		<b>07/09/2021 18:44:31 GMT</b>

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
103.000	13	BX	7513400_SAMPLE	Diversey Clearklens IPA 900 mL	FMP21175	WO52072/STR002199
			302_0112	REF: 13.5A3 AF Min = 0.717		
107.000	333	BX	101105062	Diversey Clearklens IPA 900ml	FMP21175	WO52072/STR002199
			302_0112	REF: 13.5A3 AF Min = 0.717		
110.000	4	BX	101105062	101105062 Sample	FMP21175	WO52072/STR002199
			302_0112	REF: 13.5A3 AF Min = 0.717		
	<b>350</b>	<b>BX</b>	<b>Total</b>			

**Quality Test Summary**

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.7 KGY	Pass	JBARRATT	07/10/2021 04:39:28 GMT
		Reason Code Test				JAMIE BARRATT	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.0 KGY	Pass	JBARRATT	07/10/2021 04:39:47 GMT
		Reason Code Test				JAMIE BARRATT	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS  
Reason: Work Order Completions

Date: 07/12/2021 10:38:38 GMT

# PHARMACEUTICAL ANALYSIS REPORT

# LUCIDEON

insight creating advantage

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Clear Klens IPA

**Sample Description:** Sample Code: Fmp21175 52072

**Lucideon Sample Number:** UK213134-21260

**Lucideon Report Number:** UK213134-21260/ETEP **Issue Number:** 1

**Date Logged:** 20-Jul-2021 **Order Number:** PO400364

**Date Reported:** 10-Aug-2021 **Date(s) of Test(s):** 04-Aug-2021 to 04-Aug-2021

### Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method


### Test Results:

The test results meet the EP/USP criteria: Yes

Result: <0.2 EU/ml

Tests carried out in accordance with cGMP

**End of Test Report**



10 AUG 21

Mrs Andrea Saunders  
**Business Support Administrator**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com

# PHARMACEUTICAL ANALYSIS REPORT

# LUCIDEON

insight creating advantage

## Flexible Medical Packaging

Unit 8  
Hightown  
White Cross Industrial Estate  
Lancaster  
LA1 4XS

FAO: Mrs.Olga Kirchner

**Report of Tests on:** Clear Klens IPA

**Sample Description:** Sample Code: Fmp21175 52072

**Lucideon Sample Number:** UK213134-21259

**Lucideon Report Number:** UK213134-21259/MFEP **Issue Number:** 1

**Date Logged:** 20-Jul-2021 **Order Number:** PO400364

**Date Reported:** 12-Aug-2021 **Date(s) of Test(s):** 28-Jul-2021 to 11-Aug-2021

**Sterility Testing**  
Membrane Filtration EP


## Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report**



12 AUG 21

Mrs Andrea Saunders  
**Business Support Administrator**

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited  
Queens Road, Penkhull  
Stoke-on-Trent  
Staffordshire ST4 7LQ

T +44 (0)1782 764428  
enquiries@lucideon.com  
www.lucideon.com