



Diversey Europe BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 14 October 2021

Product Name	CLEARKLENS CLEANSINALD SS VH9S		
Product Code	100848254		
Batch Number	FMP21285	52376	
Production Date	12/10/2021		
Expiration Date	EXP 12/04/2023		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		Clear Slightly Yellow Liquid
pH (neat solution)	DM001	9.0	11.4	10.8
Specific Gravity (20°C)	DM004	0.990	1.010	1.000

On behalf of Diversey site Quality Manager	Name :	Angelika Partynska Urszula Haraburda
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 02	Date of Issuing : November 24 th 2017
--------------	--------------	--

STERIS: Gamma Certificate Of Processing

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID: 2173-22774A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
C/SINALD SS 900ml Bottle DV4673	52603	13	Case

Validation Reference Number: 4673

Processing Run Start Date: 01-Oct-2021 5:01 AM

Processing Run End Date: 01-Oct-2021 11:05 PM

Specified Dose Range (kGy):	25.0 - 40.0	Calculated Min Dose (kGy):	30.0
Reference Dose Range (kGy):	27.3 - 36.6	Calculated Max Dose (kGy):	36.9

PO Number: 400685

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Gamma Process Run Approval authorized by STERIS

Date/Time E-Signed: 02-Oct-2021 2:43 AM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK
Limited, a STERIS Company
Brunel Close
Drayton Fields Industrial Estate
Daventry
Northants
NN11 8RB
Phone: + 44(0) 1327 706 111

Items irradiated under WO52603 will be used in finished batch FMP21285 52376.

Olga Kirchner

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: Cleansinald SS

Sample Description: Sample Code: Fmp21285 52376

Lucideon Sample Number: UK214442-29567

Lucideon Report Number: UK214442-29567/MFEP **Issue Number:** 1

Date Logged: 20-Oct-2021 **Order Number:** 401092

Date Reported: 09-Nov-2021 **Date(s) of Test(s):** 21-Oct-2021 to 04-Nov-2021

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

NcBoot 09-Nov-21

Mrs Natalie Boot

Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com