



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 21-09-2021

Product Name	DI ClearKlens DryClean 2x5L D*
Product Description	55% (v/v) n-propanol
Product Code	101105752
Batch Number	MUL 492384
Manufacturing Date	13-09-2021
Exp. Date / Shelf Life	13-09-2022

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

The Manufacturing/Expiration dates refer to packing date of the bulk product. As the packaging process could last for more than 1 day, a difference between the dates mentioned on this document and those printed on delivered stock is possible. For stock management purposes it is recommended that the dates printed on the packs are used.

Test	Test Method	Limits	Results
		Lower - Upper	
Appearance / Colour	MF01	Clear colorless liquid	Ok
Odour	MF02	Alcohol	Ok
pH (neat)	MF09 / DM-001	6,0 - 8,0	7,0
Relative Density (20°C)	MF08 / DM-004	0,910 - 0,914	0,910
Refractive index (20°C)	MF52 / DM-005	1,3680 - 1,3692	1,3680
Purity**	based on USP<197U>	conforms with Standard	conforms
Identity**	based on USP<197A>	conforms with Standard	conforms
Conductivity $\mu\text{S}/\text{cm}^{**}$	based on USP<645>	<3	0,30

*Product was manufactured from n-propanol (tested according to Ph. Eur.) and aqua purificata (Ph. Eur.).

**External results

Manufacturer: Multifill BV Constructieweg 25A 3641 SB Mijdrecht The Netherlands	Name :	J. Koppenol
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

OOS COA Template	Version : 04	Date of issuing : March 11 th 2019
------------------	--------------	---