



Diversey Europe Operations BV
Maarssenbroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 06 December 2021

Product Name	CLEARKLENS DE VH29		
Product Code	101101081		
Batch Number	FMP21319	53411	
Production Date	15/11/2021		
Expiration Date	EXP 15/11/2023		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Colourless Liquid		Clear Colourless Liquid
Specific Gravity (20°C)	DM004	0.880	0.888	0.887

On behalf of Diversey site Quality Manger	Name:	Angelika Partynska Justyna Staron
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version : 02 Date of issuing : November 24th

R55480102

11/29/2021 16:08:39 GMT
Page 1 of 1

Customer Name:	FLEXIBLE MED. GBP UK LANCASTER	Processing Facility:	Markham Vale	Work Order #	3011410
P.O.#	401342			Sales Order #	2919103
	25 - 45 kGy	Gamma Irradiation	Irradiation Date/Time:		11/28/2021 05:42:59 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	707	BX	101101081	Diversey Clearklens DE 900ml	FMP21319	53411/STR002096
			302_0112	RF: 13.5A3, AF Min: 0.717		
102.000	3	BX	101101081_SAMPLE	Diversey Clearklens DE 900ml	FMP21319	53411/STR002096
			302_0112	RF: 13.5A3, AF Min: 0.717		
103.000	190	BX	101105062	Diversey Clearklens IPA 900ml	FMP21322	53409/STR002199
			302_0112	RF: 13.5A3, AF Min: 0.717		
	900	BX	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.4 KGY	Pass	SWARD	11/29/2021 02:45:57 GMT
		Reason Code Test				SIMON WARD	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	39.8 KGY	Pass	ESILLS	11/29/2021 13:29:13 GMT
		Reason Code Test				EMMA SILLS	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: EMMA SILLS
Reason: Work Order Completions

Date: 11/29/2021 16:08:17 GMT

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens DE

Sample Description: Sample Code: Fmp21319 53411

Lucideon Sample Number: UK215217-34056

Lucideon Report Number: UK215217-34056/ETEP **Issue Number:** 1

Date Logged: 06-Dec-2021 **Order Number:** 401444

Date Reported: 21-Dec-2021 **Date(s) of Test(s):** 16-Dec-2021 to 16-Dec-2021

Endotoxin Test (EP)

2.6.14 Bacterial Endotoxins - Method C - Turbidimetric Kinetic Method

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass (<0.2000 EU/ml)

Tests carried out in accordance with cGMP

End of Test Report

Natalie Boot 21-Dec-21

Mrs Natalie Boot

Senior Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com

PHARMACEUTICAL ANALYSIS REPORT

LUCIDEON

insight creating advantage

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs.Olga Kirchner

Report of Tests on: ClearKlens DE

Sample Description: Sample Code: Fmp21319 53411

Lucideon Sample Number: UK215217-34055

Lucideon Report Number: UK215217-34055/MFEP **Issue Number:** 1

Date Logged: 06-Dec-2021 **Order Number:** 401444

Date Reported: 22-Dec-2021 **Date(s) of Test(s):** 08-Dec-2021 to 22-Dec-2021

Sterility Testing
Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Cont 22-Dec-21
Charlotte Arnold

Client Technical Support

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Queens Road, Penkhull
Stoke-on-Trent
Staffordshire ST4 7LQ

T +44 (0)1782 764428
enquiries@lucideon.com
www.lucideon.com