

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

#### **CERTIFICATE OF ANALYSIS**

Date 01 March 2023

Product Name	CLEARKLENS BI-SPO	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26 first off		
Product Code	7514852			
Batch Number	FMP23060	56284		
<b>Production Date</b>	01/03/2023			
<b>Expiration Date</b>	EXP 01/03/2025			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi	its	Results
		Lower	- Upper	
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	9.0	12.5	11.0
Specific Gravity (20°C)	DM004	1.004	1.020	1.006
Potential Chlorine Dioxide		100	180	138.375

On behalf of Diversey site Quality Manger	Name:	Justyna Staron Edyta Rodrígues
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version: 04	Date of issuing: March 11th 2019
1 CONTINUIATE	V CISIOII . U-	Daic of issuing . March I illi 2017



Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

#### **CERTIFICATE OF ANALYSIS**

Date 01 March 2023

Product Name	CLEARKLENS BI-SPORE BASE SOLUTION VH26 first off		
Product Code	7514852		
Batch Number	FMP23060 56284		
Production Date	01/03/2023		
Expiration Date	EXP 01/03/2025		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	1.5	2.5	1.9
Specific Gravity (20°C)	DM004	1.010	1.030	1.019

On behalf of Diversey site Quality Manger	Name:	Justyna Staron	Edyta Rodrígues
	Position	Quality (	Control Inspector

This document being issued electronically does not bear a signature

COA Template		Date of issuing: March 11th 2019
	Version : 04	

# STERIS: Gamma Certificate Of Processing

Prepared for:

**FLEXIBLE MEDICAL PACKAGING LTD (8245)** 

Gamma Process Run ID

2173-34939A

**Product Code** 

**Lot Number** 

Quantity **UOM** 

**BI-SP B3 Components DV5857** 

56310

31 Case

Validation Reference Number:

5857

Processing Run Start Date: 11-Feb-2023 7:33 PM

Processing Run End Date: 12-Feb-2023 1:30 AM

Specified Dose Range (kGy):

25.0 - 40.0

Calculated Min Dose (kGy):

28.6

Reference Dose Range (kGy):

26.1 - 35.8

Calculated Max Dose (kGy):

34.3

PO Number: 404534

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

### Gamma Process Run Approval authorized by STERIS

Date/Time E-Signed: 12-Feb-2023 8:16 AM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

#### **Processing Location**

Synergy Health Sterilisation UK Limited, a STERIS Company **Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB** Phone: + 44(0) 1327 706 111

Items irradiated under WO56310 will be used in finished batch Fmp23060 56284.

Olga Kirchner

Document ID: 13542

## PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: Fmp23060 56284

Lucideon Sample Number: UK23892-5981

Lucideon Report Number: UK23892-5981/MFEP

Issue Number:

Date Logged:

09-Mar-2023

**Order Number:** 

404772

**Date Reported:** 

29-Mar-2023

Date(s) of Test(s):

14-Mar-2023 to 28-Mar-2023

**Sterility Testing** 

Membrane Filtration EP

**Test Results:** 

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report** 

NOBOOK 29 MOT 23

Mrs Natalie Boot

Senior Business Support Administrator

## PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging

Unit 8 Hightown White Cross Industrial Estate Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: Fmp23060 56284

Lucideon Sample Number: UK23892-5982

Lucideon Report Number: UK23892-5982/MFEP

Issue Number:

Date Logged:

09-Mar-2023

**Order Number:** 

404772

Date Reported:

29-Mar-2023

Date(s) of Test(s):

14-Mar-2023 to 28-Mar-2023

**Sterility Testing** 

Membrane Filtration EP

Test Results:

The test results meet the EP/USP criteria: Yes

Result: Pass

Tests carried out in accordance with cGMP

**End of Test Report** 

NEGOOK 29 MOT 23

Mrs Natalie Boot

Senior Business Support Administrator