

Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

#### **CERTIFICATE OF ANALYSIS**

Date: 28 March 2023

<b>Product Name</b>	CLEARKLENS BI-SP	CLEARKLENS BI-SPORE ACTIVATOR SOLUTION VH26 first off		
Product Code	7514852			
Batch Number	FMP23087	56714		
<b>Production Date</b>	28/03/2023			
<b>Expiration Date</b>	EXP 28/03/2025			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi		Results
		Lower	- Upper	
Appearance	Visual	Clear Colou	ırless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	9.0	12.5	10.98
Specific Gravity (20°C)	DM004	1.004	1.020	1.018
Potential Chlorine Dioxide		100	180	131.625

On behalf of Diversey site Quality Manger	Name:	Edyta Rodrígues Justyna Staron
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template Version: 04 Date of issuing: March 11th 2019



Diversey Europe Operations BV Maarssenbroeksedijk 2 3542 DN Utrecht The Netherlands

Tel. +31 (0)30 247 6427 www.Diversey.com

### **CERTIFICATE OF ANALYSIS**

Date: 28 March 2023

Product Name	CLEARKLENS B	CLEARKLENS BI-SPORE BASE SOLUTION VH26 first off		
Product Code	7514852			
Batch Number	FMP23087	56714		
<b>Production Date</b>	28/03/2023			
<b>Expiration Date</b>	EXP 28/03/2025			

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limi Lower	ts - Upper	Results
Appearance	Visual	Clear Colou	rless Liquid	Clear Colourless Liquid
pH (neat solution)	DM001	1.5	2.5	1.9
Specific Gravity (20°C)	DM004	1.010	1.030	1.016

On behalf of Diversey site Quality Manger	Name:	Justyna Staron Edyta Rodrígues
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 04	Date of issuing: March 11th 2019
--------------	--------------	----------------------------------

## **STERIS: Gamma Certificate Of Processing**

Prepared for: FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-35885A

Product CodeLot NumberQuantityUOMBI-SP B3 Components DV58575673724Case

Validation Reference Number: 5857

Processing Run Start Date: 17-Mar-2023 3:43 AM

Processing Run End Date: 17-Mar-2023 10:05 AM

Specified Dose Range (kGy): 25.0 - 40.0 Calculated Min Dose (kGy): 30.2

Reference Dose Range (kGy): 26.1 - 35.8 Calculated Max Dose (kGy): 36.5

PO Number: 404819

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

#### Gamma Process Run Approval authorized by STERIS

Date/Time E-Signed: 17-Mar-2023 12:36 PM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

#### **Processing Location**

Synergy Health Sterilisation UK Limited, a STERIS Company Brunel Close Drayton Fields Industrial Estate Daventry Northants NN11 8RB

Phone: + 44(0) 1327 706 111

Items irradiated under WO56737 will be used in finished batch Fmp23087 56714.

Olga Kirchner

Document ID: 138414

N/A Last Revised in Rel 2.0.0.0 Rel Date: 13-Aug-2018 Page 1 of 1

## PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging Unit 8 Hightown

White Cross Industrial Estate

Lancaster LA14XS

FAO: Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Activator

Sample Description:

Sample Code: Fmp23087 56714

Lucideon Sample Number: UK231282-8729

Lucideon Report Number: UK231282-8729/MFEP

Issue Number:

Date Logged:

11-Apr-2023

**Order Number:** 

**Sterility Testing** Membrane Filtration EP 405003

13-Apr-2023 to 27-Apr-2023

**Date Reported:** 

03-May-2023

Date(s) of Test(s):

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report** 

NOBOOK O3 May 23

Mrs Natalie Boot

Senior Business Support Administrator

# PHARMACEUTICAL ANALYSIS REPORT



Flexible Medical Packaging Unit 8 Hightown White Cross Industrial Estate

Lancaster LA14XS

FAO:

Mrs.Olga Kirchner

Report of Tests on:

Bi-Spore Base

Sample Description:

Sample Code: Fmp23087 56714

Lucideon Sample Number: UK231282-8730

Lucideon Report Number: UK231282-8730/MFEP

Issue Number:

1

Date Logged:

11-Apr-2023

Order Number:

405003

**Date Reported:** 

03-May-2023

Date(s) of Test(s):

13-Apr-2023 to 27-Apr-2023

**Sterility Testing** 

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

**End of Test Report** 

