



Diversey Europe Operations BV
Maarssebroeksedijk 2
3542 DN Utrecht
The Netherlands

Tel. +31 (0)30 247 6427
www.Diversey.com

CERTIFICATE OF ANALYSIS

Date: 02 April 2025

Product Name	CLEARKLENS CLEANSINALD SS VH9S		
Product Code	100848254		
Batch Number	FMP25091	61891	
Production Date	01/04/2025		
Expiration Date	EXP 01/10/2026		

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

Test	Test Method	Limits		Results
		Lower	Upper	
Appearance	Visual	Clear Slightly Yellow Liquid		Clear Slightly Yellow Liquid
pH (neat solution)	DM001	9.0	11.4	11.0
Specific Gravity (20°C)	DM004	0.990	1.010	0.996

On behalf of Diversey site Quality Manger	Name:	Edyta Rodrigues Natalia Traichel
	Position	Quality Control Inspector

This document being issued electronically does not bear a signature

COA Template	Version : 04	Date of issuing : March 11th 2019
--------------	--------------	-----------------------------------

STERIS: Gamma Certificate Of Processing

Prepared For FLEXIBLE MEDICAL PACKAGING LTD (8245)

Gamma Process Run ID 2173-54613A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
C/SINALD SS 900ml Bottle DV4673	61896 - STR001906	25	Case
Validation Reference Number: 4673			

Processing Run Start Date 07-Mar-2025 5:58 PM

Processing Run End Date 08-Mar-2025 12:21 AM

Specified Dose Range (kGy)	25.0 - 40.0	Calculated Min Dose (kGy)	30.0
Reference Dose Range (kGy)	27.3 - 36.6	Calculated Max Dose (kGy)	37.1

PO Number: 410188

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Gamma Process Run Approval authorized by STERIS

DateTime Esigned 08-Mar-2025 6:40 AM

Operating facilities are in compliance with applicable regulations providing services under a certified quality system which meets the requirements of FDA QSR, EN/ISO 13485 current certified version, and is in alignment with EN ISO 11137 current certified version

Processing Location

Synergy Health Sterilisation UK
Limited, a STERIS Company
Brunel Close
Drayton Fields Industrial Estate
Daventry
Northants
NN11 8RB
Phone: + 44(0) 1327 706 111

Items irradiated under WO61896 will be used in finished batch FMP25091 61891.

Olga Kirchner

PHARMACEUTICAL ANALYSIS REPORT

Flexible Medical Packaging

Unit 8
Hightown
White Cross Industrial Estate
Lancaster
LA1 4XS

FAO: Mrs. Olga Kirchner

Report of Tests on: Cleansinald SS

Sample Description: Sample Code: Fmp25091 61891

Lucideon Sample Number: N251579 - 1

Lucideon Report Number: N251579 - 1/MFEP

Issue Number: 1

Date Logged: 11-Apr-2025

Order Number: 410442

Date Reported: 30-Apr-2025

Date(s) of Test(s): 15-Apr-2025 to 29-Apr-2025

Sterility Testing

Membrane Filtration EP

Test Results:

Result: Pass

The test results meet the EP/USP criteria: Yes

Tests carried out in accordance with cGMP

End of Test Report

Signed by:

Tracy Wheeler

Signer Name: Tracy Wheeler

Signing Reason: I have reviewed this document

Signing Time: 30 April 2025 | 10:35 BST

D85AB0246E184F31821E2EC1AFA6CA90

Tracy Wheeler

Business Support Administrator

Page 1 of 1

This report is issued in accordance with the Conditions of Business of Lucideon Limited and relates only to the sample(s) tested. No responsibility is taken for the accuracy of the sampling unless this is done under our own supervision. This report shall not be reproduced in part without the written approval of Lucideon Limited, nor used in any way as to lead to misrepresentation of the results or their implications.

Lucideon is the trading name of Lucideon Limited. Registered in England No. 1960455.

Lucideon Limited
Stone Business Park
Brooms Road, Stone
Staffordshire, ST15 0SH, UK
T +44 (0)1785 331874
query@lucideon.com
www.lucideon.com
Reg. England 1960455

Certificate Of Completion

Envelope Id: 767DDE32-0349-4D21-BE8B-F6AB80B963EF

Status: Completed

Subject: N251579 Flexible Medical

Source Envelope:

Document Pages: 1

Signatures: 1

Envelope Originator:

Certificate Pages: 1

Initials: 0

Tracy Wheeler

AutoNav: Enabled

Queens Road

Envelopeld Stamping: Disabled

Stoke-on-Trent, Staffordshire ST4 7LQ

Time Zone: (UTC) Dublin, Edinburgh, Lisbon, London

Tracy.Wheeler@uk.lucideon.com

IP Address: 86.188.245.148

Record Tracking

Status: Original

Holder: Tracy Wheeler

Location: DocuSign

30-Apr-25 | 10:34

Tracy.Wheeler@uk.lucideon.com

Signer Events

Signature

Timestamp

Tracy Wheeler

Sent: 30-Apr-25 | 10:35

tracy.wheeler@uk.lucideon.com

Viewed: 30-Apr-25 | 10:35

Placeholder

Signed: 30-Apr-25 | 10:36

Lucideon Limited

Signature Adoption: Pre-selected Style

Security Level: Email, Account Authentication (Required), Authentication, Login with SSO

Signature ID:

D85AB024-6E18-4F31-821E-2EC1AFA6CA90

Using IP Address: 86.188.245.148

With Signing Authentication via Docusign password

With Signing Reasons (on each tab):

I have reviewed this document

Electronic Record and Signature Disclosure:

Not Offered via Docusign

In Person Signer Events

Signature

Timestamp

Editor Delivery Events

Status

Timestamp

Agent Delivery Events

Status

Timestamp

Intermediary Delivery Events

Status

Timestamp

Certified Delivery Events

Status

Timestamp

Carbon Copy Events

Status

Timestamp

Witness Events

Signature

Timestamp

Notary Events

Signature

Timestamp

Envelope Summary Events

Status

Timestamps

Envelope Sent

Hashed/Encrypted

30-Apr-25 | 10:35

Certified Delivered

Security Checked

30-Apr-25 | 10:35

Signing Complete

Security Checked

30-Apr-25 | 10:36

Completed

Security Checked

30-Apr-25 | 10:36

Payment Events

Status

Timestamps